



California Medical Device Recall Information



Recall Name

Medtronic Recalls Shiley Neonatal and Pediatric Tracheostomy Tubes Due to the Possibility that Switching Tubes May Cause Breathing Issues

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|--------------|---|--|---|
| 5/08/15 | <ul style="list-style-type: none">Shiley Neonatal, Pediatric and Long Pediatric Tracheostomy TubeCuffless Shiley Neonatal, Pediatric and Long Pediatric Tracheostomy Tube with TaberGuard Cuff | Medtronic Navigation, Inc. North Haven, CT | <i>Potential for patients who switched from previous Shiley tracheostomy tubes to the newer model tubes to experience airway obstruction.</i> |
| Recall Class | Product Identification | Distribution | Affected Dates |
| I | List of Affected Product Numbers and Lots | CA, nationwide | Manufacturing and Distribution Dates: May 2, 2013 to April 27, 2015 |

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm454968.htm>